

# TEMPERATURE LOG

Protect Your Vaccines. Check Temperatures Twice a Day!

Facility Name \_\_\_\_\_

Mo/Yr \_\_\_\_\_

PIN \_\_\_\_\_

DAYS 1-15

Instructions: Insert the day, time, and your initials in the designated boxes. Mark an "X" in the box that corresponds to your temperature reading. Write an "N" or "Y" for both refrigerator and freezer to verify that you reviewed the alarm status on the temperature monitoring device(s). An "N" means that there is no alarm and a "Y" means that the word "alarm" is displayed on the temperature monitoring device screen. If you mark "Y" or if you record a temperature in the shaded zone, follow the instructions below.

**THE SHADED ZONES REPRESENT UNACCEPTABLE TEMPERATURES!**

*If a temperature recorded here is unacceptable or if there is an alarm status on the temperature monitoring device:*

1. Store the vaccine under proper conditions as quickly as possible. (Segregate the vaccine and identify it as "DO NOT USE" while vaccine viability is being determined).
2. Immediately download and review the graph.
3. Contact the **Epidemiology Vaccine Depot at 907-341-2202** to determine whether the potency of the vaccine has been affected.

Day of Month	1		2		3		4		5		6		7		8		9		10		11		12		13		14		15		
Day of Week																															
Time of Day	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	
Staff Initials																															
Alarm N or Y																															
REFRIGERATOR (F°)	≥48°	<b>DO NOT USE vaccines that have been stored at these temperatures! See instructions at top of page.</b>																													
	47°																														
	46°																														
	45°																														
	44°																														
	43°																														
	42°																														
	41°																														
	40°																														
	39°																														
	38°																														
	37°																														
	36°																														
	35°																														
	34°	<b>DO NOT USE vaccines that have been stored at these temperatures! See instructions at top of page.</b>																													
≤33°																															
Alarm N or Y																															
FREEZER (F°)	≥8°	<b>Note: Frost-free freezers will rise above 5°F for a short period of time (maximum of 4 times/day). This is normal and acceptable for vaccine storage.</b>																													
	7°																														
	6°	<b>Write the actual temperature in the box instead of an "X" if the temperature is below 3°.</b>																													
	5°																														
	4°																														
≤3°																															



# TEMPERATURE LOG

Protect Your Vaccines. Check Temperatures Twice a Day!

Facility Name \_\_\_\_\_

Mo/Yr \_\_\_\_\_

PIN \_\_\_\_\_

DAYS 16-31

Instructions: Insert the day, time, and your initials in the designated boxes. Mark an "X" in the box that corresponds to your temperature reading. Write an "N" or "Y" for both refrigerator and freezer to verify that you reviewed the alarm status on the temperature monitoring device(s). An "N" means that there is no alarm and a "Y" means that the word "alarm" is displayed on the temperature monitoring device screen. If you mark "Y" or if you record a temperature in the shaded zone, follow the instructions below.

**THE SHADED ZONES REPRESENT UNACCEPTABLE TEMPERATURES!** *If a temperature recorded here is unacceptable or if there is an alarm status on the temperature monitoring device:*

1. Store the vaccine under proper conditions as quickly as possible. (Segregate the vaccine and identify it as "DO NOT USE" while vaccine viability is being determined).
2. Immediately download and review the graph.
3. Contact the **Epidemiology Vaccine Depot at 907-341-2202** to determine whether the potency of the vaccine has been affected.


Day of Month	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Day of Week																
Time of Day	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm
Staff Initials																
Alarm N or Y																
REFRIGERATOR (F°)	<b>DO NOT USE vaccines that have been stored at these temperatures! See instructions at top of page.</b>															
≥48°																
47°																
46°																
45°																
44°																
43°																
42°																
41°																
40°																
39°	<b>DO NOT USE vaccines that have been stored at these temperatures! See instructions at top of page.</b>															
38°																
37°																
36°																
35°																
34°																
33°																
32°																
31°																
30°																
Alarm N or Y																
FREEZER (F°)	<b>Note: Frost-free freezers will rise above 5°F for a short period of time (maximum of 4 times/day). This is normal and acceptable for vaccine storage. Write the actual temperature in the box instead of an "X" if the temperature is below 3°.</b>															
≥8°																
7°																
6°																
5°																
4°																
3°																

Facility Name \_\_\_\_\_

PIN \_\_\_\_\_

Refrigerator/Freezer Location \_\_\_\_\_

*(maintain this form with this unit)*



## Action Taken

*(because temperatures reached or were approaching unacceptable range)*

Date	Time <small>(include a.m. or p.m.)</small>	Brief Description of Problem	Action Taken